

Software and Instrumentation Review and Cybersecurity Considerations

Lisa Simone, Ph.D.

Division of Emerging and Transfusion Transmitted Diseases

Office of Blood Research and Review

CBER

July 15, 2019


Presentation Outline

- Systems level approach to review
- Baseline topics for a risk-based software and instrumentation review
- Newer topics for challenges in the use environment
 - Interoperability
 - Cybersecurity
- Updates for Premarket Cybersecurity Guidance (in progress)

Overview of Device Configurations (1)




- Systems often include several parts to meet the intended use
- Different configurations, different workflows
- May interface to different networks
- Some systems are straightforward - all parts from the same Medical Device Manufacturer (MDM)



Overview of Device Configurations (2)



- Other systems are more complex, with multiple parts and connections
- Regulatory requirements may be different for each part
- Premarket review considers how each contributes to the risk of the overall system
- For review, additional “system level” documentation may be needed to demonstrate how all parts together are reasonably safe and effective
- Interoperability becomes important



Baseline for Software and Instrumentation Review

Guidance for Industry and FDA Staff

Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices

Document issued on: May 11, 2008

This document supersedes: Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices, issued May 29, 1998, and Reviewer Guidance for a Premarket Notification Submission for Blood Establishment Computer Software, issued January 13, 1997.

For questions regarding this document concerning devices regulated by CDER contact Linda Chomander.

Documentation for review outlined in many guidance documents

- Most familiar: “*Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices*”

Because devices operate in increasingly complex use environments:

- Cybersecurity guidance (premarket 2014 and postmarket 2016)
- Interoperability guidance (2017)

Documentation Type	Present	Adequate (Yes/No/Assessment Incomplete)
1. Level of Concern:	<input type="checkbox"/>	Choose an item.
2. Software Description:	<input type="checkbox"/>	Choose an item.
3. Device Hazard Analysis:	<input type="checkbox"/>	Choose an item.
4. Software Requirements Specifications:	<input type="checkbox"/>	Choose an item.
5. Architecture Design Chart:	<input type="checkbox"/>	Choose an item.
6. Software Design Specifications:	<input type="checkbox"/>	Choose an item.
7. Traceability Analysis/Matrix:	<input type="checkbox"/>	Choose an item.
8. Software Development Environment:	<input type="checkbox"/>	Choose an item.
9. Verification & Validation Testing:	<input type="checkbox"/>	Choose an item.
10. Revision Level History:	<input type="checkbox"/>	Choose an item.
11. Unresolved Anomalies:	<input type="checkbox"/>	Choose an item.

12. Cybersecurity:	<input type="checkbox"/>	Choose an item.
13. Interoperability:	<input type="checkbox"/>	Choose an item.

“Guidance for Industry and FDA Staff - Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices”

<https://www.fda.gov/media/73065/download>.

“Content of Premarket Submissions for Management of Cybersecurity in Medical Devices – Guidance for Industry and Food and Drug Administration Staff”

<https://www.fda.gov/media/86174/download>.

“Postmarket Management of Cybersecurity in Medical Devices - Guidance for Industry and Food and Drug Administration Staff”

<https://www.fda.gov/media/95862/download>.

“Design Considerations and Pre-market Submission Recommendations for Interoperable Medical Devices” <https://www.fda.gov/media/95636/download>.

Issues with Submissions

Documentation

Issues	Impact
Cover letter does not describe true reason for the submission	Hard to identify specific changes that are focus the review
Documentation difficult to search, hyperlinks missing or incorrect	Check links and PDF creation to allow least burdensome review. Use helpful filenames.
Complex tables rendered into microscopic PDF font sizes	Review PDFs for readability

How the software guidance documentation is used in review

Review goal: focus on what can go wrong in the system, and identify what has been done to reduce those risks to acceptable levels

- Not a “checklist” review – focus on risks
- Establish reasonable assurance of safety and effectiveness
- Reduce burden for applicant and reviewer

The documentation should support a risk-based review

Guidance for Industry and FDA Staff

Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices

Document issued on: May 11, 2008

This document supersedes: Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices, issued May 29, 1998, and Reviewer Guidance for a Premarket Notification Submission for Blood Establishment Computer Software, issued January 13, 1997.

For questions regarding this document concerning devices regulated by CDREH contact Linda Shomaker at (240) 276-6055. For questions regarding this document concerning devices regulated by CDREH contact Linda Wei at (301) 827-0138.



CIR

U.S. Department of Health and Human Services
Food and Drug Administration
Center for Devices and Radiological Health
Office of Device Evaluation
Office of In Vitro Diagnostics

Documentation Type	Present	Adequate (Yes/No/ Assessment Incomplete)
1. Level of Concern:	<input type="checkbox"/>	Choose an item.
2. Software Description:	<input type="checkbox"/>	Choose an item.
3. Device Hazard Analysis:	<input type="checkbox"/>	Choose an item.
4. Software Requirements Specifications:	<input type="checkbox"/>	Choose an item.
5. Architecture Design Chart:	<input type="checkbox"/>	Choose an item.
6. Software Design Specifications:	<input type="checkbox"/>	Choose an item.
7. Traceability Analysis/Matrix:	<input type="checkbox"/>	Choose an item.
8. Software Development Environment:	<input type="checkbox"/>	Choose an item.
9. Verification & Validation Testing:	<input type="checkbox"/>	Choose an item.
10. Revision Level History:	<input type="checkbox"/>	Choose an item.
11. Unresolved Anomalies:	<input type="checkbox"/>	Choose an item.

Risk-Based Review (1)

1. Is the system doing the right thing?



- Does it satisfy its medical purpose?

2. Is the system not doing the wrong thing?



- Does it detect and prevent error situations that could cause incorrect operation?

This is a critical aspect of a risk-based review

Risk-Based Review (2)

Approach

1. Learn about system/device

What does the device do?



Documentation Referenced

- Intended Use
- Software Description
- Device Description in submission
- Manuals, etc.

How does the device do it?

Technology behind the operation



- Usually spread throughout the submission, but focuses on:
- Architecture,
- Requirements,
- Specifications, and
- Many non-software sections


Risk-Based Review (3)

Questions Asked


2. Risk of harm

- Identify issues that have higher estimates of risk (focus on the harms that can occur)
- Identify specific risk control measures that reduce the risk of harm
- Identify testing to demonstrate risk control measures verified
- Review higher residual risks

Documentation Referenced



3. Other sources of risk








Issues with Submissions

Device Hazard Analysis

Issues	Impact
Risk management process not provided or explained	Can't evaluate if residual risks are acceptable. Align with industry standard process.
Estimations of risk prior to risk control/mitigation not provided	Can't identify which controls are the most important in reducing risk
Clear trace between individual risk control measures and their verification test missing (incomplete traceability)	Can't link risk controls to requirements and testing to determine if a risk control measure is reasonable and is verified to reduce risk
Risk analysis limited to only some parts of the overall system	Can't draw conclusions about safety and effectiveness of entire system. Provide system level analysis.
Impact of any hardware changes from previous submissions or during trials is not discussed	Device HA is not limited to software hazards

Issues with Submissions

Assay Hazard Analysis

Issues	Impact
Specific assay-related hazards/harms not included	 Can't tell if analysis considered assay-specific hazards and individual hazardous situations (e.g., typical, worst case)
MAUDE adverse event data used as an estimate of probability	 Helps identify hazard causes and contributing factors. It shouldn't be used to estimate probabilities (data quality issues, underreporting).
Risk acceptability criteria not provided, or individual benefit/risk justifications missing (if needed)	 Can't evaluate if residual risks are acceptable. Benefit/risk determinations for individual risks may be necessary.
Factors that are outside the manufacturer's control are used to reduce estimates of risk in the device hazard analysis	 Factors such as viral inactivation or presence of disease-treating drugs inform benefit/risk discussions. These are not device risk control measures.
Software was upgraded during preclinical/clinical studies	 Changes are possible if risk assessment shows no impact on the data previously collected (use pre-sub pathway for questions)

Issues with Submissions

Testing

Issues	Impact
Test plans/protocols missing	May needed to evaluate type of testing performed, esp. for higher risks
Missing verification of information for safety	Risk reduction that relies on labeling should be verified (e.g., usability tests)
Failed tests not explained/justified	Can't determine impact of failed tests on safety and effectiveness. Provide assessment.

Issues with Submissions

Other Issues

Issues

Unresolved anomalies don't include impact on safety and effectiveness, operator usage and human factors

Unclear how end user notified of anomaly-related risks/workarounds

Full documentation not provided for standalone software/SaMD



Impact

Can't determine the impact of leaving defects unresolved in marketed system. Provide justification.

Can't assess how residual risks are disclosed to user. Include traces to labeling where risks are disclosed.


Review must consider risks related to use of all software in the system

Operating in the Use Environment: Interoperability and Cybersecurity



Interoperability: Two or more products, technologies or systems exchanging and using information

- Example information exchanged:
 - patient data
 - assay information
 - instrument data
 - command and control over other devices
 - mobile notifications, etc.
- Example purposes:
 - support intended use
 - receive software updates
 - perform backup/restore
 - service/maintenance, etc.
- Connectivity leads to increased risks



The Cybersecurity Environment



- Software in connected medical devices is vulnerable to threats
- Cybersecurity incidents can directly impact medical devices or networked operations
- When vulnerabilities are not addressed, malware might enter and spread through user, lab, hospital/healthcare facility networks
- Compromise of data confidentiality, integrity, and availability may lead to patient harm, through:
 - Compromise of critical device functionality
 - Delay in diagnosis/treatment intervention

Interoperability Guidance:

- List on externally-facing electronic interfaces (EIs)
 - Purpose, role and anticipated users
 - Impact on device performance
 - How the interface is used, and limitations
- Risk analysis including security-related issues
- V&V under normal and abnormal conditions that are reasonably likely to occur
- Information in Labeling

Cybersecurity Guidance:

- Hazard analysis related to intentional and unintentional cybersecurity risk
- Traceability matrix linking cybersecurity risks to controls
- Summary of plan to provide validated software updates and patches
- Summary of controls to assure medical device will maintain its integrity throughout design and release
- Recommended cybersecurity controls (e.g., antivirus, firewall)

Contains Nonbinding Recommendations

Design Considerations and Pre-market Submission Recommendations for Interoperable Medical Devices

Guidance for Industry and Food and Drug Administration Staff

Document issued on: September 6, 2017
The draft of this document was issued on January 26, 2016.

For questions about this document regarding CDRH-regulated devices, email them to:
DigitalHealth@fda.hhs.gov.

For questions about this document regarding CBER-regulated devices, contact the Office of Communication, Outreach and Development (OCOD), by calling 1-800-835-4709 or 240-402-8010.

Content of Premarket Submissions for Management of Cybersecurity in Medical Devices

Guidance for Industry and Food and Drug Administration Staff

Document Issued on: October 2, 2014

The draft of this document was issued on June 14, 2013.

For questions regarding this document contact the Office of Device Evaluation at 301-796-5550 or Office of Communication, Outreach and Development (CBER) at 1-800-835-4709 or 240-402-7800.



U.S. Department of Health and Human Services
Food and Drug Administration
Center for Devices and Radiological Health
Office of Device Evaluation
Office of In Vitro Diagnostics and Radiological Health
Center for Biologics Evaluation and Research

Issues with Submissions

Interoperability

Issues

Assumption that interoperability guidance does not apply

List of externally-facing electronic interfaces (EIs) not provided, although hardware may list multiple USBs, network ports, etc.

Connected functionality mentioned without reference to EIs, protocols, protections

Impact

Guidance drives risk identification for all functionality used as part of the system. Analysis should include unintentional misuse and malicious use.

Prevents clear picture of connectivity and risks associated with connected functionality. List/discuss EIs individually.

Hinders review of risk and cybersecurity considerations. Include requested information.

Issues with Submissions

Cybersecurity

Issues

- ➡ Diagrams of system components not provided (e.g., network diagrams, data flow, etc.)
- ➡ Cybersecurity controls not linked to specific cybersecurity risks
- ➡ Cybersecurity is treated like a silo
- ➡ Not assuming the worse case scenario for a security-related risk
- ➡ Not hardening the system to prevent access of unused ports

Issues

- ➡ Not disclosing residual risks to users to inform their risk management activities (shared responsibility)
- ➡ Not considering End of Support dates for operating systems
- ➡ Not considering the security risks associated with use of off-the-shelf (OTS) software, and therefore not validating OTS software for security, in addition to safety and effectiveness

FDA Cybersecurity History

FDA




Vulnerabilities and Medical Devices



- FDA's thinking has evolved based on external security events and our review of documentation provided
- Recent issues drive FDA's actions
 - Hackable infusion pump that may over- or under-infuse drugs
 - Implanted cardiac devices that might stop working or work incorrectly
- Exploits might not target medical devices directly
 - WannaCry impacted those who hadn't installed a security update for Windows XP (legacy issue)
- Issues on the horizon
 - Ransomware in the short term
 - Attacks that influence the physics of sensors to change their input and outputs; e.g., using radio waves, acoustics
 - Tampering of medical records and trustworthiness of chart data used to treat and diagnose patients

Recalled Device for Vulnerability in Off-the-Shelf Software



Possible harm associated
with incorrect patient results

Updated Cybersecurity Premarket Guidance: What's New



- Because medical device cybersecurity continues to evolve, new guidance is needed
- Designing trustworthy devices – security spans the entire product lifecycle
 - Integrating threat modeling
 - Secure development lifecycle
 - Considering “exploitability” of a vulnerability rather than estimating probabilities
 - Software Bill of Materials
- Shifting the mindset to scenarios in the use environment beyond “intended use”
- Engage in proactive behavior and information sharing
- Preventing multi-patient (i.e., scaled) attacks

Parting Thoughts for Software and Instrumentation Review

Documentation Needs

Play your part in a least burdensome review

- Review applicable guidance documents for what to provide
- Provide great risk analysis to guide a risk-based review
- Ensure V&V covers higher risks at a minimum
- Anticipate reviewer questions: proactively explain any discrepancies, failed tests, anomalies, use of multiple software versions for testing, etc.

Our Documentation Asks

- Documentation must support review of the entire system for
 - safety and effectiveness, and
 - security
- If system contains elements from more than one manufacturer, agreements might be necessary to allow FDA to review the necessary documentation
- When in doubt, use FDA's presubmission pathway for advice



Thank you!

Lisa Simone
lisa.simone@fda.hhs.gov

Guidance Reference List

The following is a list of the most common guidances considered when designing medical devices and for establishing documentation to support premarket reviews. This is not an exhaustive list.

- “Guidance for Industry and FDA Staff - Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices,” issued May 11, 2005 and available at <https://www.fda.gov/media/73065/download>.
- “Guidance for Industry and FDA Staff: “Assay Migration Studies for In Vitro Diagnostic Devices,” issued April 25, 2013 and available at <https://www.fda.gov/media/73669/download>.
- “Content of Premarket Submissions for Management of Cybersecurity in Medical Devices – Guidance for Industry and Food and Drug Administration Staff,” issued October 2, 2014 and available at <https://www.fda.gov/media/86174/download>.
- “Postmarket Management of Cybersecurity in Medical Devices - Guidance for Industry and Food and Drug Administration Staff,” issued December 28, 2016 and available at <https://www.fda.gov/media/95862/download>.
- “Guidance for Industry, FDA Reviewers and Compliance on Off-The-Shelf Software Use in Medical Devices,” issued September 9, 1999 and available at <https://www.fda.gov/media/72154/download>.
- “General Principles of Software Validation; Final Guidance for Industry and FDA Staff,” issued January 11, 2002 and available at <https://www.fda.gov/media/73141/download>.
- “Guidance for Industry - Cybersecurity for Networked Medical Devices Containing Off-the-Shelf (OTS) Software,” issued January 14, 2005 and available at <https://www.fda.gov/media/72154/download>.
- “Design Considerations and Pre-market Submission Recommendations for Interoperable Medical Devices,” issued September 5, 2017 and available at <https://www.fda.gov/media/95636/download>.
- “Radio Frequency Wireless Technology in Medical Devices - Guidance for Industry and Food and Drug Administration Staff,” issued August 14, 2013 and available at <https://www.fda.gov/media/71975/download>.
- “Design Control Guidance for Medical Device Manufacturers,” issued March 11, 1997 and available at <https://www.fda.gov/media/116573/download>.

Cybersecurity References

- FDA's Website on Cybersecurity: <https://www.fda.gov/medical-devices/digital-health/cybersecurity>
 - Mitigating Cybersecurity Risks
 - Cybersecurity Guidelines
 - Cybersecurity Safety Communications
 - Reporting Cybersecurity Issues
 - MOUs on Cybersecurity in Medical Devices
 - Workshops and Webinars on Cybersecurity
 - Other Collaborations on Cybersecurity
 - Cybersecurity in the News
- Medical Device Safety Action Plan (April 2018): <https://www.fda.gov/about-fda/cdrh-reports/medical-device-safety-action-plan-protecting-patients-promoting-public-health>
- AAMI BI&T: The Evolving State of Medical Device Cybersecurity March/April 2018: <https://www.aami-bit.org/doi/full/10.2345/0899-8205-52.2.103>
- Perspective piece in American Heart Association Journal 'Circulation' (Sept 2018:): <https://www.ahajournals.org/doi/10.1161/CIRCULATIONAHA.118.035021>
- Report on Advancing Coordinated Vulnerability Disclosure – MDIC publication (Oct 2018): <http://mdic.org/wp-content/uploads/2018/10/MDIC-CybersecurityReport.pdf>
- Suzanne B. Schwartz, MD, MBAFDA, Center for Devices and Radiological Health, USENIX 2018, Baltimore Maryland, Aug 17, 2018
- Seth D Carmody, PHD, HCISPP, CDRH / FDA, International Council on Systems Engineering Conference, May 1, 2019